

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Custom Calibration, Inc.**  
 154 State STE 106  
 North Haven, CT 06473  
 Kevin Mastriano 203-484-3707

**CALIBRATION**

Valid to: **April 17, 2025**

Certificate Number: **AC-2671**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup> (50/60 Hz)	Up to 0.5 A (0.5 to 4) A (4 to 40) A (40 to 100) A (100 to 200) A (200 to 400) A	17 mA 73 mA 0.73 A 2.1 A 5.5 A 19 A	Extech 380947 AC/DC Clamp Meter
DC Current – Measure <sup>1</sup>	(0 to 5) A (5 to 20) A (20 to 40) A (40 to 100) A (100 to 150) A (150 to 200) A (200 to 400) A	82 mA 0.26 A 0.49 A 1.4 A 2 A 5.4 A 19 A	Extech 380947 AC/DC Clamp Meter
Electrical Simulation of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω (-200 to -150) °C (-150 to 360) °C (360 to 740) °C (740 to 850) °C Pt 385, 1 000 Ω (-200 to 170) °C (170 to 470) °C (470 to 730) °C (730 to 850) °C	0.045 °C 0.072 °C 0.14 °C 0.16 °C 0.047 °C 0.089 °C 0.14 °C 0.16 °C	PIE 525B RTD and Thermocouple Calibrator



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
Electrical Simulation of Thermocouple Indicating Devices <sup>1</sup>	Type J		PIE 525B RTD and Thermocouple Calibrator	
		(-200 to -150) °C		0.21 °C
		(-150 to -50) °C		0.22 °C
		(-50 to 300) °C		0.25 °C
		(300 to 850) °C		0.3 °C
		(850 to 1 200) °C		0.36 °C
	Type K			
		(-230 to -100) °C		0.26 °C
		(-100 to 600) °C		0.32 °C
		(600 to 1 000) °C		0.35 °C
		(1 000 to 1 371) °C		0.4 °C
	Type R			
		(-18 to 250) °C		1.7 °C
		(250 to 750) °C		1.8 °C
		(750 to 1 600) °C		2 °C
		(1 600 to 1 767) °C		2 °C
	Type S			
	(-18 to 150) °C	1.7 °C		
	(150 to 500) °C	1.7 °C		
	(500 to 1 650) °C	1.9 °C		
	(1 650 to 1 767) °C	2 °C		
Type T				
	(-260 to -240) °C	0.26 °C		
	(-240 to -210) °C	0.27 °C		
	(-210 to -100) °C	0.26 °C		
	(-100 to 50) °C	0.27 °C		
	(50 to 100) °C	0.32 °C		

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1</sup>	Up to 6 in	300 μin	Direct comparison to Gage Blocks
	(6 to 12) in	310 μin	
	(12 to 18) in	320 μin	
	(18 to 24) in	340 μin	
Depth Gauges <sup>1</sup>	Up to 2 in	33 μin	Direct comparison to Gage Blocks
	(2 to 6) in	51 μin	
	(6 to 12) in	91 μin	

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dial and Test Indicators <sup>1</sup>	Up to 2 in (2 to 6) in (6 to 12) in	160 μin 160 μin 180 μin	Direct comparison to Gage Blocks
Pin Gauges <sup>1</sup> (Diameter)	Up to 1 in	25 μin	Laser Micrometer
Laser Micrometers <sup>1</sup>	Up to 1 in	36 μin	Comparison to Master Pin Gauges
Height Gauges <sup>1</sup>	Up to 6 in (6 to 12) in (12 to 18) in (18 to 24) in	51 μin 91 μin 140 μin 180 μin	Direct comparison to Gage Blocks
Length Standards and Rods <sup>1</sup>	(0.1 to 40) in	990 μin	Comparison to Standard Rule
Micrometers <sup>1</sup>	Up to 2 in (2 to 6) in (6 to 12) in	34 μin 51 μin 140 μin	Direct comparison to Gage Blocks
Rules <sup>1</sup>	Up to 1 000 mm	12 μm	Direct Comparison to Standard Rule
	Up to 40 in	990 μin	
Thickness Gauges <sup>1</sup>	Up to 6 in (6 to 12) in (12 to 18) in (18 to 24) in	90 μin 81 μin 130 μin 170 μin	Direct comparison to Gage Blocks

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances <sup>1,2</sup>	(0.001 to 120) g (120 to 200) g (220 to 5 000) g	2.2 mg 4 mg 90 mg	ASTM E617 Class 1 Weights and internal procedure CM-WHT01 utilized in the calibration of the weighing system.
Scales and Balances <sup>1,2</sup>	(1 to 50) lb (50 to 100) lb (100 to 300) lb (300 to 500) lb	0.005 8 lb 0.012 lb 0.035 lb 0.066 lb	NIST Class F Weights and internal procedure CM-WHT01 utilized in the calibration of the weighing system.



ANSI National Accreditation Board

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force Device <sup>1</sup>	Up to 100 lbf (100 to 200) lbf	0.21 lbf 0.4 lbf	Direct Comparison to Load Cell
Torque Measuring Devices <sup>1</sup>	Up to 90 lbf·in	0.58 lbf·in	HIOS HP100 Digital Torque Meter
Pressure Devices <sup>1</sup>	Up to 15 psig (15 to 30) psig (30 to 100) psig (100 to 300) psig (300 to 500) psig (500 to 10 000) psig	0.018 psi 0.023 psi 0.072 psi 0.22 psi 0.36 psi 7.2 psi	Direct Comparison to Master Pressure Gauges
Vacuum Devices <sup>1</sup>	(-15 to 0) psiv	0.053 psi	Direct Comparison to Master Vacuum Gauges

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-100 to 200) °C	0.19 °C	4132 Platinum RTD Thermometer
	(200 to 500) °C	2.4 °C	Type K Thermocouple Probe w/ Indicator
	(500 to 1 000) °C	4.7 °C	
Relative Humidity – Measure <sup>1</sup>	(10 to 95) %RH	2.7 %RH	Extech RH390 Precision Psychrometer

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RPM Measure – Contact <sup>1,3</sup>	(0.5 to 500) rpm (500 to 1 000) rpm (1 000 to 2 000) rpm (2 000 to 3 000) rpm (3 000 to 4 000) rpm	0.42 rpm 0.96 rpm 2 rpm 2.7 rpm 3.1 rpm	Extech 461995 Contact/Photo Tachometer


**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RPM Measure – Non-Contact <sup>1,3</sup>	(1 to 60) rpm (60 to 1 000) rpm (1 000 to 2 000) rpm (2 000 to 3 000) rpm (3 000 to 4 000) rpm (4 000 to 5 000) rpm (5 000 to 10 000) rpm (10 000 to 20 000) rpm (20 000 to 50 000) rpm (50 000 to 99 999) rpm	0.58 rpm 1.6 rpm 2.1 rpm 2.5 rpm 3.2 rpm 4.3 rpm 6.7 rpm 13 rpm 42 rpm 65 rpm	Extech 461995 Contact/Photo Tachometer
Linear Distance Speed – Contact <sup>1,3</sup>	(5 to 50) ft/min (50 to 100) ft/min (100 to 200) ft/min (200 to 500) ft/min (500 to 1 000) ft/min (1 000 to 2 000) ft/min (2 000 to 3 000) ft/min (3 000 to 4 000) ft/min	0.74 ft/min 1.3 ft/min 2.5 ft/min 5.9 ft/min 12 ft/min 24 ft/min 35 ft/min 47 ft/min	Extech 461995 Contact/Photo Tachometer
Timers and Stopwatches	60 s to 24 hr	0.42 s	Comparison to Control Company 1025 Digital Stopwatch

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The CMC presented here does not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
3. rpm = revolutions per minute.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2671.



Jason Stine, Vice President